



Industrie Service

The ATP Testing and Calibration Station
of TÜV SÜD Industrie Service GmbH

confirm to the company

Pointer Telocation Ltd.
7715 NW 48th Street
Suite 395 Doral
FL 33166 USA

because of the positive results of the tested built-in thermometer,

MultiSense
MultiSense-TH

according to the test-report No. **1129 TM** from the 22nd of January 2018, that the built-in thermometer fulfils the requirements of the standard DIN EN 13485 for the determination:

Gerät	Multisense	Multisense-TH
Genauigkeitsklasse 0,5 in °C	---	-40 bis 85
Genauigkeitsklasse 1 in °C	-40 bis 40	---
Genauigkeitsklasse 2 in °C	40 bis 85	---
Umgebungsbedingung	E	E

The manufacturer is allowed to mark the built-in thermometer with the following test-report number:

ATP-MUC 1129 TM

This certificate
is valid from 22.01.2018
This certificate
is valid to: 21.01.2023

Munich, the 22nd of January 2018

Center of competence
Refrigeration and air-conditioning

Andreas Klotz



The test engineer

Thorsten Bennert